



# PERRY JOHNSON LABORATORY ACCREDITATION, INC.

## *Certificate of Accreditation*

*Perry Johnson Laboratory Accreditation, Inc. has assessed the Laboratory of:*

***Statim Lab - Division of Orum International S.r.l.***  
*Via Novara, 89, Milano (MI), Italy 20153*

*(Hereinafter called the Organization) and hereby declares that Organization is accredited in accordance with the recognized International Standard:*

**ISO/IEC 17025:2017**

This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (as outlined by the joint ISO-ILAC-IAF Communiqué dated April 2017):

***Mechanical Calibration***  
***(As detailed in the supplement)***

Accreditation claims for such testing and/or calibration services shall only be made from addresses referenced within this certificate. This Accreditation is granted subject to the system rules governing the Accreditation referred to above, and the Organization hereby covenants with the Accreditation body's duty to observe and comply with the said rules.

For PJLA:

Tracy Szerszen  
President

*Initial Accreditation Date:*

February 14, 2023

*Issue Date:*

February 14, 2023

*Expiration Date:*

May 31, 2025

*Revision Date:*

February 23, 2024

*Accreditation No.:*

120810

*Certificate No.:*

L23-116-R1

Perry Johnson Laboratory  
Accreditation, Inc. (PJLA)  
755 W. Big Beaver, Suite 1325  
Troy, Michigan 48084

*The validity of this certificate is maintained through ongoing assessments based on a continuous accreditation cycle. The validity of this certificate should be confirmed through the PJLA website: [www.pjllabs.com](http://www.pjllabs.com)*



# Certificate of Accreditation: Supplement

## Statim Lab - Division of Orum International S.r.l.

Via Novara, 89, Milano (MI), Italy 20153

Contact Name : Mrs. Alessia Ligugnana Phone: 024-875-2467

Accreditation is granted to the facility to perform the following calibrations:

### Mechanical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	CALIBRATION OR MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY ( $\pm$ )	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED	CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED
Air Flow Sampler <sup>FO</sup> (Fixed Points)	25 L/min	0.29 % of reading	IM-RM G16, Rotary Gas Meter with accessories	Internal Calibration procedure no. 7021 (Air Flow sampler)
	50 L/min	0.32 % of reading		
	100 L/min	0.53 % of reading		
	180 L/min	0.51 % of reading		
	200 L/min	0.99 % of reading		
Flow Rate - Measure <sup>FO</sup> Air Flowmeters (Fixed Points)	25 L/min	0.33 % of reading		Internal Calibration procedure no. 7022 (Flow Rate Measure)
	50 L/min	0.47 % of reading		
	100 L/min	1.56 % of reading		
	150 L/min	0.39 % of reading		
	200 L/min	1.04 % of reading		
	250 L/min	1.07 % of reading		

1. The CMC (Calibration and Measurement Capability) stated for calibrations included on this scope of accreditation represents the smallest measurement uncertainty attainable by the laboratory when performing a more or less routine calibration of a nearly ideal device under nearly ideal conditions. It is typically expressed at a confidence level of 95 % using a coverage factor  $k$  (usually equal to 2). The actual measurement uncertainty associated with a specific calibration performed by the laboratory will typically be larger than the CMC for the same calibration since capability and performance of the device being calibrated and the conditions related to the calibration may reasonably be expected to deviate from ideal to some degree.
2. The presence of a superscript F means that the laboratory performs calibration of the indicated parameter at its fixed location.
3. The presence of a superscript O means that the laboratory performs calibration of the indicated parameter onsite at customer locations
4. Measurement uncertainties obtained for calibrations performed at customer sites can be expected to be larger than the measurement uncertainties obtained at the laboratories fixed location for similar calibrations. This is due to the effects of transportation of the standards and equipment and upon environmental conditions at the customer site which are typically not controlled as closely as at the laboratories fixed location.